

**Rajasthan Medical Services Corporation Limited, Jaipur**

**Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005**

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : [edepmrmisc-rj@nic.in](mailto:edepmrmisc-rj@nic.in)

CIN:U24232RJ2011SGC035067

Website: <http://rmisc.health.rajasthan.gov.in>

No.:F-8( ) RMSC/EPM/M-3/NIB-592/2021-22/ 7955

Dated: 2/8/2021

**Rate Contract and Price Details**

S.N	Item	Detail
1	Name of Equipment / Item	Oxygen Concentrator (10 LPM)
2	Name & Address of Firm	M/S Mann Electronics, E-55, Road No. 3, Indraprastha Industrial Area, Kota, India-324005
3	Indicative Qty	1000
4	Status of Bidder	Manufacturer
5	Performance Security	425000.00(Four lac twenty five thousand only)
6	Brand/ Make	Mann
	Model	MN-1024-10 A Dual
7	Packing Unit	Each
8	HSN Code	90192090
9	R/C No. & Validity	R.C. No. 680 (Valid Upto 30.6.2023)
10	<b>Rates for supply to RMSC from Suppliers</b>	
	(i) New Basic Rate (in Rs. )	75893.00
	(ii) IGST (in Rs. )	0.00
	(iii) CGST @ 2.5%(in Rs. )	1897.00
	(iv) SGST @ 2.5%(in Rs. )	1897.00
	<b>(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs. )</b>	<b>79687.00</b>
11	<b>Rate For Supply to Consignee From RMSC</b>	
	(i) Basic Rate (in Rs.)	75893.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	3795.00
	(iii) CGST@ 2.5% { On (i)+(ii)} (in Rs.)	1992.00
	(iv) SGST@ 2.5% { On (i)+(ii)} (in Rs.)	1992.00
	<b>(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)</b>	<b>83672.00</b>
13	<b>Technical Specification of Oxygen Concentrator (10 LPM)</b>	
	1. Should have O2 (two) flowmeters having flow from 0-5 litres per minute with variation of 0.5LPM.	
	2. It should have provision of connecting 2 patients at a time.	
	3. Purity of oxygen should be more than 93%.	
	4. Suspension Type Oil free Air Compressor.	
	5. Dual Flow meter 0-5 LPM/5-10 LPM.	



6. Should have high temperature machine should shut off in case of high temperature.
7. Should have audio visual alarm for power failure, high/low oxygen level.
8. Should have oxygen purity Indicator and alarm if oxygen purity level decreases below 85%.
9. Should have built in castors for easy mobility and a handle/handgrip for easy movement facility.
10. Should have digital LED/LCD timer showing aggregate run time of the machine and also time counter since machine put on for use.
11. Should have Inbuilt box for keeping accessories like mask/cannula/power cord/cover and manual etc.
12. O2 air suction inlet filters should be there.
13. Power supply 220-240 VAC, 50Hz.
14. Should have in built safety mechanism for voltage fluctuation.
15. Should be able to function on inverter/generator in case of power failure.
16. Should be user friendly, quite (for restful sleep) & safe.
17. Should be provided with cover, cannula/mask & humidifier-2 sets, 5 set of filter and descriptive working manual.
18. Accessories should be supplied as per 2 flowmeters.
19. Should be provided with 3 years warranty.
20. High-efficiency Lithium Molecular Sieve.
21. Self-Diagnostic Function with Multiple Alarm Systems.
22. 9-Layer Filtration System to get pure Oxygen.
23. The noise level should be less than 50 db.
24. Preferably USFDA/European CE with notified body/BIS.
25. Weight should be less than 25 Kg.
26. Rates of accessories should be quoted separately.

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

  
Executive Director (EPM)  
RMSCL, Jaipur

Dated: 2/8/2021

No.:F-8() RMSC/EPM/M-3/NIB-592/2021-22/ 7955

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Jaipuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. BME (R&M), RMSC, Jaipur.
12. M/S Mann Electronics, Kota
13. Guard file.

  
Executive Director (EPM)  
RMSCL, Jaipur

**Rajasthan Medical Services Corporation Limited, Jaipur**

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrmisc-rj@nic.in

CIN : U24232RJ2011SGC035067

Website : www.rmssc.nic.in

No.:F-8( ) RMSC/EPM/M-3/NIB-592/2021-22/ 7954

Dated: 2/8/2021

All Principals Medical Colleges,  
All the Superintendents of Medical Colleges & Attached Hospitals.  
Director (PH/HA/IEC/AIDS/RCH)

**Sub:- Approval/Purchase of Item, 'Oxygen Concentrator 10 LPM', for the Rate Contract and supply, ending on 30.06.2023.**

Please find enclosed herewith approved firm and rates etc. in respect of item 'Oxygen Concentrator 10 LPM', for the rate contract and supply ending on 30.06.2023. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website [www.rmssc.nic.in](http://www.rmssc.nic.in).

In reference to the above rate contract No. F-8( ) RMSC/EPM/M-3/NIB-592 /2021-22..7955..Dated 2/8/2021.... the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.

  
Executive Director (EPM)  
RMSCL, Jaipur

No.:F-8( ) RMSC/EPM/M-3/NIB-592/2021-22/7954

Dated: 2/8/2021

Copy forwarded to the following for information &amp; necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to PHS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. P.S. to Mission Director NHM Jaipur.
4. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
5. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
6. Principal R.U.H.S. College of Dental Science Jaipur,
7. Principal, RUHS College of Medical Sciences, Jaipur.
8. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
9. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
10. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
11. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
12. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
13. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
14. All P.M.O.'s.
15. All CM&HO's.
16. All Dy. CM&HO's.
17. AGM (IT), RMSC, Jaipur.
18. M/S Mann Electronics, Kota
19. Guard file.

  
Executive Director (EPM)  
RMSCL, Jaipur