

**Rajasthan Medical Services Corporation Limited, Jaipur**  
**Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005**

Ph. No. 0141-2223887, Fax : 0141-2228065  
 CIN:U24232RJ2011SGC035067

E-mail : [edepmrmisc-ri@nlc.in](mailto:edepmrmisc-ri@nlc.in)  
 Website: <http://rmisc.health.rajasthan.gov.in>

No. F-8(JRMSC/EPM/M-4//NIB-527/2020-21/1124

Dated 28/09/2021

**Rate Contract and Price Details**

S.N.	Item	Detail
1.	Name of Equipment / Item	Syringe Needle Destroyer.
2.	Name & Address of Firm	M/s United Surgical Industries, 445, F.I.E. Patparganj Industrial Area, Delhi Email : <a href="mailto:info@unisurg.com">info@unisurg.com</a>
3.	Bidder Status	Manufacturere
4.	Performance Security	BG No. 2129IGP002231821 Dated 03-09-2021 for Rs. 1,15,000 of Bank of Baroda Patparganj Branch, Valid upto 30-11-2024
5.	Brand/Model	UNITED/USI-806
	HSN Code	90185090
6.	Packing Unit	Each
7.	R/C No. & Validity	R.C. No 691 (Valid Up to 30 Sep 2023)
8.	Indicative Quantity	4000
9.	<b>Rates for supply to RMSC from Suppliers</b>	
	(i) Basic Rate (in Rs.)	1006.00
	(ii) IGST @ 12% (in Rs.)	120.72
	(iii) CGST (in Rs.)	0.00
	(iv) SGST(in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	1126.72
10.	<b>Rate For Supply to Consignee From RMSC</b>	
	(i) Basic Rate (in Rs.)	1006.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	50.30
	(iii) CGST@ 6% { On (i)+(ii)} (in Rs.)	63.37
	(iv) SGST@ 6% { On (i)+(ii)} (in Rs.)	63.37
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	1183.04
11.	<b>Technical Specifications of Syringe &amp; Needle Destroyer</b>	
	1. Should be lightweight, portable and compact and safe destruction of sharps and syringes. 2. Should provide a removable and reusable discharge tray made for easy disposal of syringe hubs. 3. Should have a high carbon steel cutter to cut syringes. Able to cut the needle along with hub of any size. And able to destroy needles of type up to 16G. 4. Housing should be moulded type, shock proof and made of ABS plastic/stainless steel (SS-304).	

*[Signature]*  
28/9/21



5. Should have a power On/Off switch and an indication for power.
6. Should be able to destroy minimum of 7 needles on continuous operation.
7. Should works on AC power supply of 220-240V/50Hz and properly insulated for the protection from electrical hazard.
8. Should provide with 5 numbers of fuse of adequate rating.
9. Needle container should have bio-hazard symbol of size 35 mm (Black/Red) on the front & back face.
10. Product should be manufactured by ISO 9001:2008 standards.
11. Test report of stainless steel-(SS-304).
12. Final technical approval after product demonstration.
13. Guarantee: Three year against any manufacturing defect.

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)  
RMSC, Jaipur

No. F-8()RMSC/EPM/M-4//NIB-527/2020-21/

Dated

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
21. M/s United Surgical Industries, 445, F.I.E. Patparganj Industrial Area, Delhi Email : info@unisurg.com
22. Guard file.

Executive Director (EPM)  
RMSC, Jaipur