



**Rajasthan Medical Services Corporation Limited, Jaipur**

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrmsc-rj@nic.in

CIN : U24232RJ2011SGC035067

Website : www.rmssc.nic.in

No.:F-8( ) RMSC/EPM/SM-3/NIB-476/500

Dated: 19/03/2021

All Principals Medical Colleges,  
All the Superintendents of Medical Colleges & Attached Hospitals.  
Director (PH/HA/IEC/AIDS/RCH)

**Sub:-** Approval/Purchase of Item, 'Deep Freezer (-80)DegreeC(Cap.250 Plasma Bags), for the Rate Contract and supply, ending on 31.03.2023.

Please find enclosed herewith approved firm and rates etc. in respect of item "Deep Freezer (-80)DegreeC(Cap.250 Plasma Bags), for the rate contract and supply ending on 31.01.2023. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website [www.rmssc.nic.in](http://www.rmssc.nic.in).

In reference to the above rate contract F-8( ) RMSC/EPM/SM-3/NIB-476/663

Dated: the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.

Executive Director (EPM)  
RMSCL, Jaipur

Date: 19/03/21

No No.: F-8( ) RMSC/EPM/SM-3/NIB-476/500

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. M/s Remi Elektrotechnik Limited, Mumbai.
22. Guard file.

Executive Director (EPM)  
RMSCL, Jaipur

**Rajasthan Medical Services Corporation Limited, Jaipur****Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur – 302005**

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrmisc-rj@nic.in

CIN : U24232RJ2011SGC035067


Website :www.rmisc.nic.in

No.: F-8() RMSC/EPM/SM-3/NIB-476/ 501

Dated: 19/08/2021

**Rate Contract and Price Details**

S.n	Item	Detail
1	Name of Equipment / Item	'Deep Freezer (-80)DegreeC(Cap.250 Plasma Bags),
2	HSN Code	<b>8418</b>
3	Name & Address of Firm	M/S Remi Elektrotechnik Limited, Remi House, 11 Cama Indl. Estate, Goregoan(E), Mumbai-400063 E-mail :- bids@remilabworld.com
4	Bidder Status	Manufacturers
5	Brand/ Make	Remi Elektrotechnik Limited
	Model	PDV 360 Ultra with Stabilizer
6	Packing Unit	Each
7	R/C No. & Validity	R.C. No 663 (Valid Upto 31.03.2023)
8	Indicative Quantity	10
9	<b>Rates for supply to RMSC from Suppliers</b>	
	(i) New Basic Rate (in Rs. )	395000
	(ii) IGST @ 18% (in Rs. )	71100
	(iii) CGST (in Rs. )	0.00
	(iv) SGST (in Rs. )	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs. )	466100
10	<b>Rate For Supply to Consignee From RMSC</b>	
	(i) Basic Rate (in Rs.)	395000
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	19750
	(iii) CGST@ 9% { On (i)+(ii)} (in Rs.)	37328
	(iv) SGST@ 9% { On (i)+(ii)} (in Rs.)	37328
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	489406



11 Technical Specifications of item (-) 80 degree C Deep freezer (Cap. 250 Plasma Bags)

1. The Deep Freezer should achieve low temperature of -80 degree C.
2. Should be heavy duty refrigeration system, maintenance free, with hermetically sealed refrigeration compressors reliable refrigeration with minimum noise & vibration.
3. Must have uninterrupted digital electronic temperature circular chart recorder with 21 days memory with print facility / six inch, 24×7-day, ink-less pressure sensitive circular chart recorder and accuracy should be +/- 1 degree Celsius with supply of free charts for a period of Guaranty.
4. Provision to connect with central (Temperature) monitoring system / facility of wireless data transfer facility & storage of temp. data on cloud, with continuous accessibility of data on PC or laptop and on smart phones with internet facility.
5. Should have a chart range of -100 degree Celsius to +50 degree Celsius.
6. Construction of double wall with efficient insulation to minimize temperature loss, inner chamber should be made of SS 304 grade non corrosive stainless steel & outer made of high quality C/R sheet.
7. Equipment shall be supplied with external Servo Voltage Stabilizer with input voltage rating 90-280V and 50 Hz.
8. Hold over time - 2 Hrs at ambient temp.
9. External ambient temp. - Can perfectly maintain internal temperature as above at full load in an ambient temperature of +10 to at least +25 °C.
10. Cooling down time –
  - A full load of plasma packs at +25 °C takes at least 5 hr for all packs to reach below -5°C.
  - A full load of plasma packs at +25 °C takes maximum 30 hrs for all packs to reach below - 20 °C.
11. Manufacturing should be compliant with ISO 13485 (Valid documentation should be submitted in technical bid).
12. Quoted model should be compliant with European CE (Notified Body) /US FDA/BIS (Valid documentation should be submitted in technical bid).
13. The Deep Freezer should be vertical upright with a capacity to hold minimum 250 plasma bags of 250 ml.
14. Should have eye level display of temperature.
15. Should have handle for opening the door.
16. Should have audio alarms for the following: Temperature High/Low, Power Failure/ Battery low, door open.
17. Should have battery back-up for monitoring of cabinet temperature for 8 hrs (for full charge) by sealed maintenance free battery.
18. Should have at least 3 Height Adjustable/Fix trays (4 Compartments).
19. Should have display of 3\*7 Segment LED/LCD display of working status of equipment at eye level.
20. Should have lockable heavy duty caster wheels.
21. Should have stainless steel trays.
22. Noise factor should not exceed 60 decibel.
23. Should have Microprocessor based digital temperature controller.
24. Should have high density polyurethane or equivalent gaskets.
25. Should have 12.5 cm foamed in PUF insulation in exterior walls.
26. Should have triple rubber gasket sealing.
27. Should have hot line around the mouth of the cabinet which prevents condensate formation.
28. Should have 4 stainless steel interior compartment doors with magnetic latches ensure storage and less cold air loss during opening and closing of the doors.
29. Should have caster wheels as standard feature.

  
17/3/21



30. Should have temperature controller with 0.1 degree Celsius with battery backup.
31. Should have rust free stainless steel door hinge makes the door open up to an angle of 135 degrees.
32. Should have well located appropriate/proper ventilation facility which helps in better condensation process.
33. Storage rack should be easily removable.
34. Rear clamps ensure the correct distance from walls, thereby ensuring good air circulation.
35. Guarantee: Three years on equipment from the date of installation.
36. CMC: CMC will be given @ 5 % of net rate - exclusive GST (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.
37. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
38. Installation will be done by supplier free of cost.
39. The service engineer should be based in Rajasthan.
40. Four Preventive Maintenance services during Guarantee and CMC period are essential.
41. Demonstration of equipment is must for final technical approval.

Other terms and condition to be covered during guarantee period by the supplier:-

- Response time- < 48 Hours after logged the complain in complain portal.
- Service hours- As per health facility schedule.
- Preventive Maintenance Schedule (PM)-As per OEM (Mention PM Schedule).
- Calibration Schedule – As per OEM (Mention Schedule).
- Up time- 95% (346 Days).
- Technical & Application Support, Demonstrations & Trainings - As & when required.
- Toll Free No-
- Life of the equipment- As per OEM.

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

**Executive Director (EPM)**  
**RMSC, Jaipur**

No. F-8() RMSC/EPM/SM-3/NIB-476/501

Dated: 19/03/2021

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/ Pr. Secy./ Secy., Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur,



PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.

13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. M/s Remi Elektrotechnik Limited, Mumbai.
22. Guard file.

19/12/14

Executive Director (EPM)  
RMSC, Jaipur