



Rajasthan Medical Services Corporation Limited, Jaipur
Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: <http://rmisc.health.rajasthan.gov.in>

No. F-8() RMSC/EPM/M-3/2024-25/NIB-816/

Dated:

Revised Rate Contract and Price Details

(As per price fall clause under Rule 29 (2) (h) of the RTPP Rules,2013 W.E.F.05.04.2024)

S.N.	Item	Detail
1.	Name of Equipment / Item	Digital Hemoglobin Meter Strips with lancets (packet of 50 Nos.)
2.	Name & Address of Firm	M/S Sensa Core Medical Instrumentation Pvt. Ltd. Plot No. 3, Export Promotion Industrial Park, Pashamylaram, Sangareddy (District), Hyderabad E-mail: info@sensacore.com GSTIN – 36AAKCS5301F1ZJ
3.	Bidder Status	Manufacturer
4.	Packing Unit	Each
5.	Make/Model	Sensa Core
6.	R/C No. & Validity	R.C. No. 852 (Valid Upto 31.05.2026)
7.	Performance Security	Bank Guarantee No. 6312NDDG00008825 dated 30.05.2024 for Rs. 8780980.00 valid up to 30.11.2029
8.	Indicative Quantity	561085 Pkt
9.	Rates for supply to RMSC from Suppliers	
	(i) New Basic Rate (in Rs.)	242.00
	(ii) IGST @ 12% (in Rs.)	29.04
	(iii) CGST @ 6% (in Rs.)	0.00
	(iv) SGST@ 6% (in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	271.04
10.	Rate For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	242.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	12.10
	(iii) CGST@ 6% { On (i)+(ii)} (in Rs.)	15.246
	(iv) SGST@ 6% { On (i)+(ii)} (in Rs.)	15.246
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	284.592



11.

Technical Specifications of the Hemoglobin Meter Strips With Lancet (Packet of 50 Nos.)

S.No.	Technical Specification	Comments
1.	Name	Digital Hemoglobinometer
2.	Clinical purpose	Direct hand held battery operated device used for haemoglobin testing in clinical setting/population-based screening. The device is intended to be used for quantitative measurement of haemoglobin in capillary, venous or arterial whole blood samples taken from forearm, upper arm, hand thigh, calf, or fingers.
3.	Should be used by	Medical officers/ Laboratory Technician/ Staff Nurse/ ANMs/ Mid-Level health providers at HWCs.
Technical Characteristics		
1.	Working principle	Reflectance Photometry/ Absorbance Photometry
2.	Parameter	Blood haemoglobin level
3.	Range of Hb estimation	0-20 gm/dl
4.	On screen patients result display	Yes (LCD Display)
5.	Maximum volume of sample required	Not more than 50µl (One full blood drop)
6.	Sample material	Capillary, venous or arterial whole blood
7.	Measuring time	Less than one minute
8.	Auto-calibration	Auto/self calibration
9.	Sensitivity	More than 80%
10.	Specificity	More than 80%
11.	Bias (Limits of agreement)	0.5 gm/dl (±1gm/dl)
Users Interface		
1.	Memory to store data	Desirable- up to 500 tests with date & time
2.	Bluetooth Connectivity	Desirable
3.	Data Transfer	Desirable- provision for data transfer to printer and PC
Physical Characteristics		
1.	Dimensions (Metric)	Not more than 15cm×10cm×20cm
2.	Weight (Grams)	Should not be more than 500 grams
3.	Mobility/Portability	Should be mobile & portable
Energy Source		
1.	Power requirements	Preferably battery operated. Should also be able to work on direct connection with electricity source (AC). The manufacturer must provide the charger & the cable for electricity power connection, wherever required by the equipment.
2.	Battery	3.7 volt Lithium-polymer rechargeable battery/1.5 volt 4 AA batteries/etc. should be able to perform up to 500 tests when full charged.
3.	Automatic shut down	On battery power the device should turn off after approximately of 5minutes of no use.
Environmental Consideration:		
1.	Working temperature	Should be able to perform in the temperature range of 10-40°C
2.	Working humidity	5-95%
3.	User's care/ Cleaning	The part of the equipment which comes in contact with blood should be easily cleanable. The factors which affect the haemoglobin estimation (ex. Haziness of lens) after repeated uses should be clearly mentioned in the manual. The cleaning material for the lens & strip should be easily available.
Accessories		
1.	Strip (Working Environment)	Disposable & stable at specified environmental (Working temperature- 10-40°C, Working humidity- 5-95%) condition.
2.	Strip (Storage Environment)	Should be stable at temperature of 25-30°C & humidity of 5-95%. Shelf life for storage should be at least one year.



3.	Strip packaging	Strip should be packed in packaging of 50Nos. in a air tight bottle/vial (50 lancet should be provided free of cost with each 50Nos. Strips),200Nos. Strips with 200 lancets and 1 Lancing pen Must be supplied free of cost with Digital Hemoglobin meter.
Quality Control		
1.	Intra-sample variation (Accuracy)	Should be less than 5%
2.	Control solution to check quality	Should be available for low, normaland high haemoglobin values. Calibration facility should be available.
3.	Certification	US FDA/ European CE(notified body)/ ISO 13485:2016 certification.
Service Support		
1.	Service support contact details	Contact details of manufacturer, supplier and local service agent to be provided
2.	Training	Free onsite training should be provided for the doctors, ANM and ASHA. At least, two trainings (one training at the time of installation and another training after six months i.e. refresher training)
3.	Warranty	3 Years

The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)
RMSC, Jaipur
Dated:

No. F-8() RMSC/EPM/M-3/2024-25/NIB-816/

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Pr. Secy/Secy, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar& Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Scinces, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota &Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur.
20. M/S Sensa Core Medical Instrumentation Pvt. Ltd, Hyderabad
21. Guard file.

Executive Director (EPM)
RMSC, Jaipur



Rajasthan Medical Services Corporation Limited, Jaipur

‘D’-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrm-sc-rj@nic.in

CIN : U24232RJ2011SGC035067

Website: <http://rm-sc.health.rajasthan.gov.in>

No. F-8() RMSC/EPM/M-3/2024-25/NIB-816/

Dated:

All Principals Medical Colleges,
All the Superintendents of Medical Colleges & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, “Digital Hemoglobin Meter Strips with lancets (packet of 50 Nos.)”, for the Rate Contract and supply, ending on 31.05.2026.

Please find enclosed herewith approved firm and “Digital Hemoglobin Meter Strips with lancets (packet of 50 Nos.)”, for the rate contract and supply ending on 31.05.2026. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rm-sc.nic.in.

In reference to the above rate contract No. F-8() RMSC/EPM/M-3/2023/NIB-816/.....Dated..... the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSC, Jaipur.

Executive Director (EPM)

RMSC, Jaipur

Dated:

No. F-8() RMSC/EPM/M-3/2024-25/NIB-816/

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon’ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Pr. Secy/Secy, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSC
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.’s.
17. All CM&HO’s.
18. All Dy. CM&HO’s.
19. AGM (IT), RMSC, Jaipur.
20. M/S Sensa Core Medical Instrumentation Pvt. Ltd, Hyderabad
21. Guard file.

Executive Director (EPM)

RMSC, Jaipur