



**Rajasthan Medical Services Corporation Limited**

Gandhi Block, SwasthyaBhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066, 2228064

Website: <http://rmsc.health.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067

E\_mail : [edprmsc@nic.in](mailto:edprmsc@nic.in)

Ref. No.:- F.02(83)/RMSCL/S&S/ RNA AND DNA EXTRACTION SYSTEM FOR REAGENT RENTAL (Automatic) Kit./NIB-18/2020/ 415

Dated: 28/10/2020

**Corrigendum - V**

**Subject:- Date extension and Amendment in Specifications**


**Ref.:- NIB No. F.02(83)/RMSCL/S&S/RNA AND DNA EXTRACTION SYSTEM FOR REAGENT RENTAL (Automatic) Kit./NIB-18/2020/352 Dated:-09.10.2020 (Technical bid opening due on dated -28.10.2020)**

S. No	Existing date / technical specification	Amended date / technical specification												
1	<p><b>F.02(83)/RMSCL/S&amp;S/RNA AND DNA EXTRACTION SYSTEM FOR REAGENT RENTAL (Automatic) Kit./NIB-18/2020/352 Dated:-09.10.2020</b></p> <table border="1"> <tr> <td>Last date and time of submission of online bids</td> <td>28.10.2020 at 3.00 PM</td> </tr> <tr> <td>EMD, Tender fees, RISL fees through challan and Physically</td> <td>28.10.2020 at 3.00 PM</td> </tr> <tr> <td>Date and time of opening of Online technical bids</td> <td>28.10.2020 at 04.00 PM</td> </tr> </table>	Last date and time of submission of online bids	28.10.2020 at 3.00 PM	EMD, Tender fees, RISL fees through challan and Physically	28.10.2020 at 3.00 PM	Date and time of opening of Online technical bids	28.10.2020 at 04.00 PM	<p><b>F.02(83)/RMSCL/S&amp;S/RNA AND DNA EXTRACTION SYSTEM FOR REAGENT RENTAL (Automatic) Kit./NIB-18/2020/352 Dated:-09.10.2020</b></p> <table border="1"> <tr> <td>Last date and time of submission of online bids</td> <td>02.11.2020 at 3.00 PM</td> </tr> <tr> <td>EMD, Tender fees, RISL fees through challan and Physically</td> <td>02.11.2020 at 3.00 PM</td> </tr> <tr> <td>Date and time of opening of Online technical bids</td> <td>02.11.2020 at 04.00 PM</td> </tr> </table>	Last date and time of submission of online bids	02.11.2020 at 3.00 PM	EMD, Tender fees, RISL fees through challan and Physically	02.11.2020 at 3.00 PM	Date and time of opening of Online technical bids	02.11.2020 at 04.00 PM
Last date and time of submission of online bids	28.10.2020 at 3.00 PM													
EMD, Tender fees, RISL fees through challan and Physically	28.10.2020 at 3.00 PM													
Date and time of opening of Online technical bids	28.10.2020 at 04.00 PM													
Last date and time of submission of online bids	02.11.2020 at 3.00 PM													
EMD, Tender fees, RISL fees through challan and Physically	02.11.2020 at 3.00 PM													
Date and time of opening of Online technical bids	02.11.2020 at 04.00 PM													
2.	<p><b><u>LIST OF SPECIFICATION OF HIGH THROUGHPUT AUTOMATIC RNA AND DNA EXTRACTION SYSTEM FOR REAGENT RENTAL S</u></b></p> <ol style="list-style-type: none"> <li>System should be capable of processing Minimum 32-96 sample in 1 Run or approx 1000 sample in 7-8 hours.</li> <li>System should be automated.</li> <li>System should be high throughput.</li> <li>System should have automated buffer dispenser/cartridge/magnetic bead design concept to minimize user intervention and pipetting steps (open system).</li> <li>System should have touch screen with pre-programmed &amp; user defined protocols.</li> <li>System should have cover to close during run of system.</li> <li>System should have sterilization facility/protocol.</li> <li>System should be CE-IVD/US-FDA Approved.</li> <li>Firm to attach user list.</li> <li>RNA recovery should be more than 95%.</li> <li>System should have cooling/heating facility if required as per kit protocol.</li> <li>System should be supplied complete with UPS of 2 hour battery backup.</li> <li>2 AC of 2 ton to be provided for 24 hour functioning.</li> <li>To be provided with suitable table.</li> <li>Firm to quote cost per test.</li> </ol>	<p><b><u>LIST OF SPECIFICATION OF HIGH THROUGHPUT AUTOMATIC RNA AND DNA EXTRACTION SYSTEM FOR REAGENT RENTAL S</u></b></p> <ol style="list-style-type: none"> <li>System should be capable of processing Minimum 32-96 sample in 1 Run or approx 1000 sample in 7-8 hours.</li> <li>System should be automated.</li> <li>System should be high throughput.</li> <li><b>System should have prefilled plates / cartridges / automated reagents dispensers to minimize user intervention and pipetting steps. It may be spin column / magnetic bead design concept (open / closed system; open means that kits of other make can also be used on the machine)</b></li> <li>System should have touch screen with pre-programmed &amp; user defined protocols.</li> <li>System should have cover to close during run of system.</li> <li>System should have sterilization facility/protocol.</li> <li>System should be CE-IVD/US-FDA Approved.</li> <li>Firm to attach user list.</li> <li>RNA recovery should be more than 95%.</li> <li>System should have cooling/heating facility if required as per kit protocol.</li> <li>System should be supplied complete with UPS of 2 hour battery backup.</li> <li>2 AC of 2 ton to be provided for 24 hour functioning.</li> <li>To be provided with suitable table.</li> <li>Firm to quote cost per test.</li> </ol>												

*Handwritten signature/initials*

**Note:-**

- It may be noted that if any further amendments are issued then a corrigendum will be published and informed.
- Rest of the terms and conditions will remain the same.

  
(Shrinivas Meena)  
Executive Director (Proc.)  
RMSCL