



**Rajasthan Medical Services Corporation, Swasthaya Bhawan, C-Scheme, Jaipur**

Phone No: 0141-2228065

Email: [rmsc@nic.in](mailto:rmsc@nic.in)

CIN:U24232RJ2011SGC035067

[www.rmhc.health.rajasthan.gov.in](http://www.rmhc.health.rajasthan.gov.in)

Ref No F.02(47)/RMSC/Proc./S&S (Non Drug)/NIB-09/2017/57

Dated: 07.02.2018

**List of Bidders Declared Responsive/Non Responsive**

(Tender No F.02(47)/RMSC/Proc./S&S (Non Drug)/NIB-09/2017/274 Dated:21.08.2017

Technical bid Opened on 08.11.2017)

The Details of Bidders Declared Responsive/Non Responsive as decided in the Purchase Committee meeting dated on 02.02.2018 headed by MD RMSCL.

S. No	Name of Bidder	Quoted Item No.	Responsive Items code no.	Non Responsive Items with reasons
1	Amkay Products Pvt. Ltd.	S-40, S-85, S-86(a), S-89(a)	S-40, S-85, S-86(a), S-89(a)	Nil
2	Angiplast Private Limited	S- 86(a), S-85, S-40, S-101, S-100	S- 86(a), S-85, S-101, S-100	S-40 Not Approved by technical expert committee because of poor handling.
3	Disposafe Health and life care Ltd.	S-89(a), S-89(b), S-100, S-101	Nil	S-89(a), S-89(b) 1. Firm has not submitted Market Standing Certificate. 2. Performance does not fulfill 15% of tender quantity requirement of S-89(a) 3. Annual Capacity and commitment quantity not written. 4. Specification is not as per tender requirement for S-89(b) S-100, S-101 1. Performance does not fulfill 15% of tender quantity requirement for S-100 2. Specification is not as per tender requirement. 3. Firm follows IS6190 which is deleted from tender specification and follows rubber material
4	Hi- Care Polymers Pvt. Ltd.	S-89(a), S-89(b), S-90(a)	S-89(a), S-89(b), S-90(a)	Nil
5	Anondita Healthcare	S-89(a), S-89(b)	S-89(a), S-89(b)	Nil
6	Jackson Care Product	S-85, S-86(a), S-99(b)	S-85, S-86(a), S-99(b)	Nil
7	K.A. Enterprises	S-99(b)	Nil	S-99(b) Not Approved by technical expert committee because not mention instructions on Packages.

(Handwritten signatures and initials)



S. No	Name of Bidder	Quoted Item No.	Responsive Items code no.	Non Responsive Items with reasons
8	La-Med Healthcare Pvt. Ltd.	S-100, S-101, S-40	Nil	S-100, S-101, S-40 Market Standing Certificate is not submitted.
9	Romsons Scientific & Surgical Industries Pvt. Ltd	S-3, S-40, S-85, S-86(a), S-97, S-100, S-101	S-3, S-40, S-97, S-100, S-101	S-85 Performance does not fulfill 15% of tender quantity requirement and sample not received  S-86(a) Not Approved by technical expert committee because grip not possible, very loose.
10	Medicare products Inc.	S-82	Nil	S-82 Not Approved by technical expert committee because not possible to skin craft.
		S-100, S-101	Nil	S-100, S-101 Firm follows IS6190 which is deleted from tender specification as it is for rubber type mask
11	Sara Healthcare Pvt. Ltd.	S-85, S-86(a)	S-85	S-86(a) Not Approved by technical expert committee because grip not possible, very loose.
12	St. Marys Rubbers Pvt. Ltd.	S-89(a), S-89(b)	Nil	S-89(a), S-89(b) Market Standing Certificate not submitted from address for which acknowledgement submitted.
13	Sutures India Pvt.Ltd.	S-89(a), S-89(b)	S-89(a), S-89(b)	Nil
		S-81	S-81	Nil
14	Vitromed Healthcare	S-40, S-100, S-101	S-40, S-100, S-101	Nil

Online Financial Bids (BoQ) of Responsive Bidders will be opened on 08.02.2018 at 3:00 P.M.

(4) 20-2

  
Executive Director (Proc.)  
RMSC